

**EVALUATION REPORT ON TECHNICAL BIDS OPENED ON 08-10-2016
FOR THE PROCUREMENT OF STOMACHER WITH BAGS, TENDER ENQUIRY NO. PC/DF(R&T)-IV/2016-17**

| Sr. # | Approved Specification of the Department | Name of the Firm & Address | Specifications/Brand/Model quoted by the Firm | Remarks/Recommendations with reasons |
|-------|--|---|---|--|
| 1. | <ul style="list-style-type: none"> • I Input voltage 220 volts • Capacity 50-400 ml • Rap Trunk Stainless steel • Time 30 to $\infty \pm 5\%$ sec • Without paddle speed $200 \pm 5\%$ RPM • Sterilized Bag Size 177 mm x 300 mm • Maximum recommended sample temperature $60^{\circ}\text{C} \pm 5\%$ • With all accessories including bag opener • Back Rack • Disposable Straws and with additional ten sets of bags • Country of origin western Europe, USA, Japan or equivalent • Two years warranty (service and parts) | Muaaz Medical Traders, Office # 22, 23 Lower Ground, Memona Centre, 1-Moj Darya Road, Lahore-54000 Pakistan. Ph: 042-37233911 | <ul style="list-style-type: none"> • Input 220 V – 2.5⁶V • 80-400 ml (circular bags) • Not given • 0 seconds – 99 minutes 59 seconds ($\pm 5\%$) • No load paddle speeds 200 rpm ($\pm 5\%$) • 177 x 305 mm • Sample temperature 60°C • Bag Opener (optional) • Back rack (optional) • Not given • Stomacher 400 circulator (Seward- UK) • Not given | The quoted specifications of this firm for stomacher with bags fulfill/qualify the standardized/approved specification. The firm is recommended for purchase of stomacher with bags subjected to provision of all accessories including back opener, rap trunk stainless steel, back rack, disposable straws and with ten sets of bags and two years warranty (service and parts) |
| 2. | <ul style="list-style-type: none"> • I Input voltage 220 volts • Capacity 50-400 ml • Rap Trunk Stainless steel • Time 30 to $\infty \pm 5\%$ sec • Without paddle speed $200 \pm 5\%$ RPM • Sterilized Bag Size 177 mm x 300 mm • Maximum recommended sample temperature $60^{\circ}\text{C} \pm 5\%$ • With all accessories including bag opener | Scientific Technical Corporation (Pvt) Ltd. LG-2, 3, 4 & 6 Queed's Center, 33-Shahrah-e-Fatima Jinnah (Queens Road) Lahore-54000 (Pakistan) | <ul style="list-style-type: none"> • 220 V, 50-60 Hz • Useful Volume: 50-400 mL • All stainless steel • Variable time (30-210 secs. Or ∞) • Adjustable paddles (patented) • Not given • Not given • Bag opener including | The quoted specifications of this firm for stomacher with bags fulfill/qualify the standardized/approved specification. The firm is recommended for purchase of stomacher with bags subjected to provision of all accessories including without paddle speed $200 \pm 5\%$ RPM, sterilized bag size 177 mm x 300 mm, maximum |

| | | | | |
|----|--|---|---|---|
| | <ul style="list-style-type: none"> • Back Rack • Disposable Straws and with additional ten sets of bags • Country of origin western Europe, USA, Japan or equivalent • Two years warranty (service and parts) | | <ul style="list-style-type: none"> • Bag rack including • Disposable straw and with 10 sets of bags • Stomacher with bags/laboratory blender, Model: Bag Mixer 400 P, Inter Science, France • 02 years from the date of supply | recommended sample temperature 60°C ± 5 % and two years warranty (service and parts) |
| 3. | <ul style="list-style-type: none"> • I Input voltage 220 volts • Capacity 50-400 ml • Rap Trunk Stainless steel • Time 30 to ∞ ± 5 % sec • Without paddle speed 200 ± 5 % RPM • Sterilized Bag Size 177 mm x 300 mm • Maximum recommended sample temperature 60°C ± 5 % • With all accessories including bag opener • Back Rack • Disposable Straws and with additional ten sets of bags • Country of origin western Europe, USA, Japan or equivalent • Two years warranty (service and parts) | Pak Glorious Enterprise, Flat # 305, Muhammadia Plaza, Gordon College Road, Rawalpindi Pakistan. Ph: +92-51-5535591 | <ul style="list-style-type: none"> • Input 220 V, 50 Hz • Capacity 80-400 ml • Not given • 0 seconds – 99 minutes 59 seconds (± 5%) • No load paddle speeds 200 rpm (± 5%) • 177 x 305 mm • Sample temperature 60°C • Bag Opener 10 qty • Back rack 10 pieces • Not given • Stomacher 400 premium (Seward- UK) • 2 year warranty and 10 year service support | The quoted specifications of this firm for stomacher with bags fulfill/qualify the standardized/approved specification. The firm is recommended for purchase of stomacher with bags subjected to provision of rap trunk stainless steel, disposable straws and with additional ten sets of bags and two years warranty with parts. |
| 4. | <ul style="list-style-type: none"> • I Input voltage 220 volts • Capacity 50-400 ml • Rap Trunk Stainless steel • Time 30 to ∞ ± 5 % sec • Without paddle speed 200 ± 5 % RPM | Saad Scientific Supplier, Suite # 13 Ground Floor, Wajid Centre Muj darya Road, Lahore-54000 Tel: 042-37324032 | <ul style="list-style-type: none"> • Input 220 V • 80-400 ml (circular bags) • Not given • 0 seconds – 99 minutes 59 seconds (± 5%) • No load paddle speeds 200 rpm (± 5%) | The quoted specifications of this firm for stomacher with bags fulfill/qualify the standardized/approved specification. The firm is recommended for purchase of stomacher with bags |

| | | | |
|--|--|---|---|
| <ul style="list-style-type: none"> • Sterilized Bag Size 177 mm x 300 mm • Maximum recommended sample temperature 60°C ± 5 % • With all accessories including bag opener • Back Rack • Disposable Straws and with additional ten sets of bags • Country of origin western Europe, USA, Japan or equivalent • Two years warranty (service and parts) | | <ul style="list-style-type: none"> • 177 x 305 mm • Sample temperature 60°C • Bag Opener (optional) • Back rack (optional) • Not given • Stomacher 400 circulator (Seward- UK) • Not given | <p>subjected to provision of all accessories including back opener, rap trunk stainless steel, back rack, disposable straws and with ten sets of bags and two years warranty (service and parts)</p> <p><i>[Signature]</i> Deputy Director Fisheries (P&D) Headquarter. Department of Fisheries PUNJAB.</p> <p><i>[Signature]</i></p> |
|--|--|---|---|

[Signature]
SENIOR CHEMIST
Fisheries Research Institute
LAHORE

[Signature]
LIBRARIAN
Fisheries Research & Training Institute,
Lahore.

[Signature]
PRINCIPAL CHEMIST
FISHERIES RESEARCH & TRAINING
INSTITUTE LAHORE.

[Signature]
Director Fisheries
Research & Training Institute
LAHORE
Research & Training Institute
Director Fisheries